

# CSIR – Central Scientific Instruments Organisation

No. CSIR-CSIO/MNOC/P07/T-10-10

Date: 18.03.2024

## 1. ICP Microphone with pre-amplifier ;Quantity 1 no

Sl .no	Parameter	Specification
01	Microphone Diameter	1/4"
02	Sensitivity	45 mV/Pa or better
03	Electrical Connector(Output)	BNC jack
04	Frequency Response	Atleast 20 to 10000 Hz
06	TEDS	Compliant

Interested firms may send their quotation through e-mail to [surjitkaman@csio.res.in](mailto:surjitkaman@csio.res.in) to the undersigned along with authorization certificate on or before 19.03.2024,12 pm

### Other Terms & Conditions:

1. Delivery: F.O.R. CSIO Chandigarh on or before 22/03/2024
2. Payment Terms: Within 30 days after delivery

Surjit Kaman

Ph:8876690733