

CSIR – Central Scientific Instruments Organisation

No. CSIR-CSIO/MNOC/HCP0101-1 /01

Date: 14.02.2024

1. Microphone with integrated pre-amplifier

Sl .no	Parameter	Specification
01	Microphone Diameter	1/4" or less
02	Sensitivity	45 mV/Pa or better
03	Electrical Connector(Output)	BNC jack /SMB/microdot
04	Frequency Response	30 Hz-15,000Hz or more *(±10 Hz accepted on lower sides
05	Type	ICCP/ICP/CCP based
06	TEDS	Compliant

and associated cable /adapter with BNC jack (if electrical output is not BNC jack).

Interested firms may send their quotation through e-mail to surjitkaman@csio.res.in to the undersigned along with authorization certificate on or before 18.02.2024

Other Terms & Conditions:

1. Delivery: F.O.R. CSIO Chandigarh
2. Payment Terms: Within 30 days after delivery

Surjit Kaman

Ph:8876690733